



### **SECTION 3 : QUESTIONS RECEIVED POST THE BRIEFFING SESSION...CONTINUATION**

#### **Question 40a**

Item: RT24-02-026ME and RT24-02-027ME

Nr: 1.8 The alternating operating cycle shall be variable up to 10 minutes (approximately)

- Does the cycle have to run for 10 minutes or should the operator be able to change the time of the cycle depending on the patient's condition?
- T1.8 - "The alternating operating cycle shall be variable up to 10 minutes (approximately)". The specification already require a cycle to be variable (able to be CHANGED).

#### **Response 40a**

The technical specification remains. If a bidder has a system that can go above 10min, please indicate on Comment section and BEC will make determination.

#### **Question 40b**

RT24-02-028ME and RT24-02-029ME Overlay

Nr 1.1 Mattress shall support patients up to 300 kg.

- No product exists to manage this safe working load. Highest patient weight available is 280kg

#### **Response 40b**

BEC will make the determination during evaluation. Comment section afford bidders to state the range of their product.

#### **Question 40c**

Nr 1.11.4 Cell height inflated: 135mm (This is too little for bariatric overlay, one need at least 20cm lift for bariatric use)

#### **Response 40c**

The technical specification remains. Bidders with a product with better technical specification can indicate on Comment section

#### **Question 40d**

RT24-02-029ME Overlay

1.11.3 Width: 850mm - Patients with weight of 300kg will not fit on a mattress width of 850mm it has to be wider.



**national treasury**

Department:  
National Treasury  
REPUBLIC OF SOUTH AFRICA

**Response 40d**

The technical specification remains. Bidders with a product with better technical spec can indicate on Comment section

Please note that no recommendation for amendment. Any deviation can be noted on the Comment section and deliberated. These items went through consultation phase with industry during Nov 2019.

---

**END**

---