

This Matrix Report is based on the Contract Circular Report									
Bid Number	RT76-1-2011ME								
Bid Description:	Supply and delivery of surgical gloves and sterilization related items to the State for the period ending 31 January 2013								
Scenario Description:	Final scenario								
Sorting Criteria:	Highest Points Awarded								
Item No	Quantity Required	Description							
		Glove Surgical Synthetic Hypo-allergenic Powder, latex and pyrogen free STERILE To comply with the latest issue of SANS 68, proof of compliance must be submitted. Must be manufactured from synthetic material Length: 283mm Size: 6 Packaging or product must state: synthetic material used Packaging to be labelled: Latex Free and Powder Free Packaging: per pair Samples must be submitted in original packaging							
RT76-01-036ME	See Quantity Awarded								
End User	Response Fields	Manta Medical							
Nat Dept of Health	Responses								
	Quantity Awarded - 21804.0000	50							
	Delivered Price	19.82							
	Total Points	91							
	Brand Name/Catalogue Number	Biogel Skinsense Size 6							
	Delivery Period (Days)	2							
	Minimum order Quantities	50							
	UOM Offered/Spec Deviation	50							
	Supplier Item Number	BG50960							
	Settlement days	30							
Item No	Quantity Required	Description							
		Tape Autoclave Indicator with adhesive To comply with the latest issue of CKS 224 and proof of compliance must be submitted Length per roll: 50 m Width: 12 mm Packaging: per roll Samples must be submitted							
RT76-04-001ME	See Quantity Awarded								
End User	Response Fields	Palmed Medical & Surgical Supplies, cc							
Nat Dept of Health	Responses								
	Quantity Awarded - 47261.0000	47261							
	Delivered Price	12							
	Total Points	95							
	Brand Name/Catalogue Number	Bastos Viegas							
	Delivery Period (Days)	30							
	Minimum order Quantities	72							
	UOM Offered/Spec Deviation	0							
	Supplier Item Number	451-003							
	Settlement days	30							

Item No	Quantity Required	Description							
		Tape Autoclave Indicator with adhesive To comply with the latest issue of CKS 224 and proof of compliance must be submitted Length per roll: 50m Width: 24 mm Packaging: per roll Samples must be submitted							
RT76-04-002ME	See Quantity Awarded								
End User	Response Fields	Palmed Medical & Surgical Supplies, cc							
Nat Dept of Health	Responses								
	Quantity Awarded - 23108.0000	23108							
	Delivered Price	22.5							
	Total Points	95							
	Brand Name/Catalogue Number	Bastos Viegas							
	Delivery Period (Days)	30							
	Minimum order Quantities	36							
	UOM Offered/Spec Deviation	0							
	Supplier Item Number	451-005							
	Settlement days	30							
Item No	Quantity Required	Description							
		Tape Masking For hospital use Crêped Type 1 To comply with the latest issue of CKS 245 and proof of compliance must be submitted Length: 50 m Width: 12 mm Packaging: per roll Samples must be submitted							
RT76-04-004ME	See Quantity Awarded								
End User	Response Fields	National Medical Supplies, cc							
Nat Dept of Health	Responses								
	Quantity Awarded - 6627.0000	6627							
	Delivered Price	9.62							
	Total Points	95.5							
	Brand Name/Catalogue Number	Wipak Steriking ITN19							
	Delivery Period (Days)	3							
	Minimum order Quantities	48							
	UOM Offered/Spec Deviation	19 mm offered and not 12 mm as required							
	Supplier Item Number	ITN19							
	Settlement days	30							
Item No	Quantity Required	Description							
		Tape Masking For hospital use Crêped Type 1 To comply with the latest issue of CKS 245 and proof of compliance must be submitted Length: 50 m Width: 24 mm Packaging: per roll Samples must be submitted							
RT76-04-005ME	See Quantity Awarded								
End User	Response Fields	Unitrade 1032, cc							
Nat Dept of Health	Responses								
	Quantity Awarded - 104957.0000	104957							
	Delivered Price	9.49							
	Total Points	96							
	Brand Name/Catalogue Number	Masking Tape							
	Delivery Period (Days)	10							
	Minimum order Quantities	50							
	UOM Offered/Spec Deviation	Roll							
	Supplier Item Number	UTB							
	Settlement days	30							

Item No	Quantity Required	Description							
		Indicator, steam sterilization Must validate the sterilization process in porous load vacuum and pulsing steam autoclaves Must be a chemical indicator The indicator must be manufactured from pure pulp board Must comply with ISO 11140-1 Proof of compliance must be submitted Samples to be submitted with original pack label 250 Indicators							
RT76-07-005ME	See Quantity Awarded								
End User	Response Fields	Qestmed (Pty) Ltd							
Nat Dept of Health	Responses								
	Quantity Awarded - 1447.0000	1447							
	Delivered Price	65.69							
	Total Points	94.6							
	Brand Name/Catalogue Number	Vista Bueno							
	Delivery Period (Days)	10							
	Minimum order Quantities	6							
	UOM Offered/Spec Deviation	n/a							
	Supplier Item Number	SAHSI001							
	Settlement days	30							
Item No	Quantity Required	Description							
		Indicator, steam sterilization Must validate the sterilization process in porous load vacuum and pulsing steam autoclaves Must be a chemical indicator The indicator must be manufactured from pure pulp board Must comply with ISO 11140-1 Proof of compliance must be submitted Samples to be submitted with original pack label 250 Indicators							
RT76-07-006ME	See Quantity Awarded								
End User	Response Fields	Qestmed (Pty) Ltd							
Nat Dept of Health	Responses								
	Quantity Awarded - 1015.0000	1015							
	Delivered Price	65.69							
	Total Points	94.6							
	Brand Name/Catalogue Number	Vista Bueno							
	Delivery Period (Days)	10							
	Minimum order Quantities	6							
	UOM Offered/Spec Deviation	n/a							
	Supplier Item Number	SAHSI 001							
	Settlement days	30							
Item No	Quantity Required	Description							
		Indicator, vacuum test pack Must test the efficiency of porous load autoclaves The pack must be made from virgin pulp paper holding in its center the chemical indicator sheet The entire pack must be strengthened by a paper wrap binding, which guarantees the constant compactness of the test pack Must have external chemical indicator Must comply with the latest edition of EN 867-4 Proof of compliance must be submitted Samples to be submitted with original pack label 20 Units							
RT76-07-007ME	See Quantity Awarded								
End User	Response Fields	Qestmed (Pty) Ltd							
Nat Dept of Health	Responses								
	Quantity Awarded - 8355.0000	8355							
	Delivered Price	413.2							
	Total Points	94.6							
	Brand Name/Catalogue Number	Vista Bueno							
	Delivery Period (Days)	10							
	Minimum order Quantities	4							
	UOM Offered/Spec Deviation	n/a							
	Supplier Item Number	SAHBD 001							
	Settlement days	30							