national treasury

Department:
National Treasury
REPUBLIC OF SOUTH AFRICA

RT24-1-2021

REVISED NOTICE 2 UPDATES

TECHNICAL SPECIFICATIONS

Please be informed that the following Category 2 series items have been added on the Annexure 2 Technical Specifications and Pricing Schedule

RT24-02-022ME	Mattress for hospital bed
RT24-02-024ME	Pressure Mattress
RT24-02-025ME	Pressure Mattress Cover
RT24-02-038ME	Infant Cot Mattress
RT24-02-039ME	Infant Cot Mattress cover
RT24-02-046ME	Mattress, for trolley/Stretcher
RT24-02-047ME	Cover, for mattress

The annexure is not locked, kindly do not tamper with any of the technical specifications wording

1. On item **RT2-002-003** Section 1.29.3 – the height adjustable dimensions have been revised from (**260 – 600mm**) to (**600-900mm**). Changes on Annexure 2 are in blue

2. On item **RT2-002-017** Section 1.3 – the highest range has been revised from (highest 500 to **800**) mm to (highest 500 to **900**) mm.

3. On item **RT2-002-017** Section 1.10 – safety sides can either be full length or half size length if bed can be in a full chair position

4. On item **RT2-002-018** Section 1.3 – the highest range has been revised from (highest 500 to **800**) mm to (highest 500 to **900**) mm. Changes on Annexure 2 are in blue

5. On item **RT2-002-018** Section 1.14 – safety sides can either be full length or half size length if bed can be in a full chair position

6. On item **RT2-002-020** Section 1.3 – the highest range has been revised from (highest 500 to **800**) mm to (highest 500 to **900**) mm. Changes on Annexure 2 are in blue

7. On item **RT2-002-020** Section 1.17 – safety sides can either be full length or half size length if bed can be in a full chair position

End