BID DESCRIPTION	SUPPLY AND DELIVERY OF HOSPITAL WARD AND THEATRE FURNITURE AND WARD REQUIREMENTS TO THE STATE FOR THE PERIOD 1 OCTOBER 2013 TO 30 SEPTEMBER 2015	
ITEM NO:	**DESCRIPTION**	**REASON**
RT24-02-022ME	Pressure relief mattress system with pump (150kg) Mattress shall support patients up to 150 kg. To comply with CKS 243. Bidder must indicate if their product complies.	Cancel: Item will be replaced by item RT24-02-023ME
RT24-04-008ME	Cabinet, utility, mobile (ABS Plastic) To serve the same function as drawing on Annexure H, drawing ME 00.30.35, see attached. ABS Material where appropriate Any mild steel sheet must be 1,2 mm thick Mild steel with epoxy/nylon power-coated finish.	No sample offered for physical evaluation:Re-tender
RT24-05-001ME	Commode, wheel chair type With bucket attachment, including bucket. To comply with the specifications in appendix F, see attached. Mild steel with epoxy/nylon powder-coated finish to comply with SANS 778 paragraph 5.2, proof of compliance must be sub	Offers received did not comply to specifications:Re-tender
RT24-07-005ME	Table, utility To comply with the specifications in Appendix P, Type B, see attached. Stainless steel table-top, without bottom shelf Stainless steel must comply wit the latest issue of SANS 417, subsection 4.1.2, proof of compliance must be submitted.	No sample offered for physical evaluation:Re-tender
RT24-07-007ME	Table, Mayo To comply with the specifications in Appendix N, see attached. Constructed entirely of stainless steel Stainless steel must comply with the latest issue of SANS 417, subsection 4.1.2, proof of compliance must be submitted. Depth of tray:	Offers received did not comply to specifications:Re-tender
RT24-08-007ME	Mid upper arm circumference (MUAC) tape, child Tape/armband to measure upper arm circumference of children Threads into itself through a hole at one end Arrow marks to indicate the point of measurement PVC, flexible, not stretchable Washable Thickn	No sample offered for physical evaluation. Re-tender
RT24-08-008ME	Mid upper arm circumference (MUAC) tape, adult Tape/armband to measure upper arm circumference of adults Threads into itself through a hole at one end Arrow marks to indicate the point of measurement PVC, flexible, not stretchable Washable Thicknes	No Quote:Re-tender
RT24-08-009ME	Stadiometer (Height board) To measure height of children and adults while standing Range: 20 - 125 cm, graduated in m, cm and mm, with numbers at metre and cm positions Measuring surface must be free of all moving parts that may affect the markings.	Offers received did not comply to specifications:Re-tender
RT24-11-001ME	Stool, anaesthetist, no castors To comply with the specification in Appendix J, Type B, see attached. Stainless steel gliders on legs Aluminium flat seat Height adjustments, floor to top of seat: 485 mm to 685 mm Spindle to be screwedwith square thre	Offers received did not comply to specifications:Re-tender

RT24-11-004ME	Stool, foot, double step To comply with the specification as per drawing ME 00.30.38 see attached. Aluminium chequered plate top Stainless steel edging Mild steel frame and legs with epoxy/nylon powder-coated finish, to comply with SANS 778.	Cancel item and revise specifications and re-tender
RT24-12-007ME	Trolley, anaesthetist. Type 5, size B. To comply with the latest issue of SANS 726 (see fig 5 and table 2). Length: 610 (± 5 mm) Width: 455 (± 5 mm) The upper part is enclosed to contain a drawer (with no divisions) opening to the front.	Offers received did not comply to specifications:Re-tender
RT24-12-024ME	Trolley, recovery, complete with mattress To comply with the latest issue of CKS 373, (fig. 2 design II to be used as a general guide) Mild steel with epoxy/nylon powder-coated finish to comply with SANS 778 paragraph 5.2, Proof of compliance must be su	Items RT24-12-024ME to RT24-12-026M are a series and bidders did not comply to the requirements of the entire series: Re-tender
RT24-12-025ME	Mattress for recovery trolley To comply with the latest issue of SANS 1291-1, proof of compliance must be submitted Must be compatible with item RT24-13-024ME To be constructed of flexible polyurethane foam complying with class 30, Grade No. 12,. Ite	Items RT24-12-024ME to RT24-12-026M are a series and bidders did not comply to the requirements of the entire series: Re-tender
RT24-12-026ME	Cover, for mattress, recovery trolley Must be compatible with item RT24-12-025ME To comply with the latest issue of SANS 1291-1, prrof of compliance must be submitted. Cover must be removable waterproof, non-slip. Items RT24-12-024ME to RT24-12-026ME	Items RT24-12-024ME to RT24-12-026M are a series and bidders did not comply to the requirements of the entire series: Re-tender
RT24-12-034ME	Carrier, soiled linen, metal box Push handle mounted on one side of unit Lockable hinged top/lid with handles Restricting mechanism to hold lid in open position (to prevent lid from hanging on hinges) Durable nylon/rubber bumper strip around the base	Items RT24-12-034ME to RT24-12-037M are a series and bidders did not bid for the entire series: Re-tender
RT24-12-035ME	Carrier, for bin, bin for soaking purposes The frame must be designed to allow easy access to the tap, effective drainage and removal of the bin Foot pedal operated plastic lid secured to frame Size: Height : 960 mm (±10 mm) Width : 400 mm (±10mm)	Items RT24-12-034ME to RT24-12-037M are a series and bidders did not bid for the entire series: Re-tender
RT24-12-036ME	Bin, for soaking purposes, removable with drainage tap (to fit carrier) Made from heat resistant nylon/plastic With handles With drainage tap Capacity : ± 85 litre Must be compatible with carrier in item RT24-12-035ME. Items RT24-12-034ME to RT24-1	Items RT24-12-034ME to RT24-12-037M are a series and bidders did not bid for the entire series: Re-tender
RT24-12-037ME	Bin, for soaking purposes, removable without drainage tap (To fit carrier) Made from heat resistant nylon/plastic With handles Without drainage tap Capacity : ± 85 litre Must be compatible with carrier in item RT24-12-035ME. Items RT24-12-034ME to RT	Items RT24-12-034ME to RT24-12-037M are a series and bidders did not bid for the entire series: Re-tender
TOTAL LINES	18	